

# iCheck Iron

## Product Information

iCheck Iron is a portable photometer for the quantitative determination of iron in foods.

### HOW DOES IT WORK?

iCheck consists of 2 parts: a ready-to-use reagent vial and a device. The sample is injected into the reagent vial where a reaction with iron takes place. The vial is inserted into the device that measures concentration of Iron in the vial.



1. Injection



2. Reaction



3. Measurement

### PRODUCT DETAILS



iCheck device

iCheck Iron measures quantitatively the iron specific colour reaction between the iron in the food and the reagents in the vials. The device displays the results of the iron content in mg/L. iCheck devices come in a portable case with all necessary accessories.



Reagent vial

iCheck reagent vials contain a patented mixture of reagents. They come in a box (Test Kit) sufficient for 100 analyses. The reagent's shelf-life is 12 months at room temperature.




### BENEFITS

- **Speed:** result in 10 to 60 minutes
- **Economy:** cost is only 10% of conventional lab. methods
- **Easy implementation:** only 1 day training required
- **Scalability:** no set up calibration required

### OUR SERVICES

#### Free-of-charge customer tech support:

- Online demos and trainings (i.e. Skype)
- Instant support via WhatsApp:  +49 1573 1111234
- Support with analysis, calculations, interpretation of standards, sampling protocols, technical consultations about the micronutrients

#### On-site Training

#### Feasibility testing for new matrices

iChecks are manufactured in Germany, used in over 80 countries and validated against standard laboratory methods. Learn more at [www.bioanalyt.com/products](http://www.bioanalyt.com/products)

#### measure for life

For orders, technical assistance or any other inquiries contact BioAnalyt at: [contact@bioanalyt.com](mailto:contact@bioanalyt.com) • +(49) 33 28 35 15 000 • [www.bioanalyt.com](http://www.bioanalyt.com)

# iCheck Iron

## Technical Data

### Quality assurance

iCheck and iCheck Iron reagent vials are produced according to quality management system (DIN EN ISO 9001:2015) certified by TÜV Nord in Germany.

| TECHNICAL DATA                               |  |
|--|--|
| Sample                                       |  |
| Analyte:                                     | Iron as NaFeEDTA, ferrous fumarate, ferrous sulfate or ferrous nitrate   |
| Sample:                                      | Premix, flour, soy and fish sauces, corn soy blend (CSB), lipid-based nutrient supplement (LNS)  |
| Sample preparation:                          | For solid samples: dilution and homogenization in distilled or bottled water, optionally in 0.2-0.4M hydrochloric acid solution.         |
| Sample volume per analysis:                  | 0.4 mL (400 µL)  |
| Concentration range:                         | >1.5 ppm (mg/L), samples above 12 ppm must be diluted in water or 0.2M HCl   |
| Device                                       |  |
| Analytical method:                           | Photometric determination of iron concentration using colorimetric reaction with bathophenanthrolin                                      |
| Units displayed:                             | mg/L   |
| Linear range:                                | 1.5 - 12.0 mg/L  |
| Calibration:                                 | Factory set (standards included for control)   |
| Time per analysis:                           | 10 - 60 min  |
| Environment:                                 | 20 -30°C, no direct sunlight   |
| Accuracy at 95% confidence interval at 25°C: | ±7-34% (depends on the sample type)  |
| Method comparison:                           | Atomic Absorption Spectroscopy (AAS)   |
| User training:                               | 1 day training   |
| Use:   | Laboratory and field   |
| Data output:                                 | Sample #, Batch #, Result, Date, Time (in transferred data)  |
| Connectivity and data:                       | Results are stored in the device and transferred to a PC via USB   |
| Power source:                                | NiMH rechargeable batteries included; 1.2V or 1.5V   |
| Warranty:                                    | 2 years  |
| Device weight:                               | 0.45 kg  |
| Device dimensions:                           | 11 x 4 x 20 cm (W x H x L)   |
| Test Kit                                     |  |
| Content:                                     | 100 reagent vials and 10 additive vials; 110 syringes - 1.0 mL; 100 needles - 1.6mm x 25mm; 10 needles - 0.8mm x 16mm; 1 syringe - 10 mL |
| Chemical composition:                        | Bathophenanthrolin in organic solvent, reducing and chelating agents   |
| Volume per reagent vial:                     | 1.5 mL   |
| Shelf life:                                  | 12 months at 20 -30°C, no direct sunlight, upright   |
| Dimension of test kit:                       | 26 x 14.5 x 16.5 cm  |
| Disposal instructions:                       | Hazardous waste  |
| Optional equipment:                          | Manual centrifuge, 50 mL falcons, weighing dishes, reference samples   |

measure for life

For orders, technical assistance or any other inquiries contact BioAnalyt at:  
contact@bioanalyt.com • +(49) 33 28 35 15 000 • www.bioanalyt.com