

Report No.: PLE230117

Test Item: Photometer

Type or Model: iCheck T3

Manufacturer: BioAnalyt GmbH
Rheinstr. 17
14513 Teltow

Arrival Date: 23. January 2023

Place of Testing: PRO EMV Labor Strausberg GmbH
Garzauer Chaussee Phone: +493341 - 3065199
15344 Strausberg Fax: +493341 - 3068441
e-mail: proemv@proemv.de www.proemv.de

Standards: EN IEC 61326-1 /2021 Immunity: table 1

Date of Testing: 25. January 2023

Procedure: The device under test was tested for compliance with the
referenced standards.
(Selected tests: see section 4)

Test Result: The test item meets the requirements.

Tested by: T. Haugk
Test Engineer

2023-01-25

Date, Signature



Inspected : R. Erxleben
General Manager

2023-01-25

Date, Signature